



BIO-SHIELD

T-2/HT-2

B3248 / B3296

ELISA test for the quantitative determination of T-2/HT-2
in grains, cereals and animal feeds.

In vitro analysis
Storage at 2-8°C

www.prognosis-biotech.com

This ELISA kit is manufactured by ProGnosis Biotech Ltd and complies with the specifications on the Standard EN ISO 14675: 2003

Prognosis Biotech is ISO 9001 certified by TÜV Hellas (TÜV NORD).

Use only the current version of Product Data Sheet enclosed with the kit.

Bio-Shield T-2/HT-2, B3248/B3296, is an immunoassay method that determines the T-2/HT-2 in grains, cereals and other commodities including animal feeds. The ELISA kit contains all reagents required for the immunoassay method. The ELISA test is adequate for 96/48 definitions (standards are included). A spectrophotometer for microtiter ELISA plate is required.

Matrices:

- **Grains:** Corn, Wheat, Barley, Rice, Oats, Oat Bran, Soy Beans
- **Other:** Sesame, Beans, Milk

- Sample preparation: extraction
- Test time (incubation time after samples and reagents preparation): 15min
- Standard curve range: 0 - 500ppb
- Shelf life: 12 months
- Storage: 2-8°C

TABLE OF CONTENTS

1. Description	4
2. General Information	4
3. Principle of the Method	4
4. Reagents Provided	4
5. Materials required but not provided	5
6. Storage Instructions	5
7. Safety and Precautions for use	5
8. Indication of corruption of kit reagents	5
9. Sample preparation	5
10. Method Procedure	6
11. Data Analysis	7
12. Example of Standard Curve	8
13. Immunoassay Specification	8
14. Performance Evaluation	9
15. Method Summary	10

1. Description

Bio-Shield T-2/HT-2 is an ELISA test for the detection of Trichothecene T-2/HT-2 in grains, cereals and animal feeds.

2. General Information

T-2 and HT-2 toxin belong to the group of trichothecenes. This group of mycotoxins are produced mainly by fungi of the genus *Fusarium* which is toxic to humans and animals. Agricultural commodities are frequently infected by this fungus. It is frequently implicated in cytotoxic and immunosuppressive disorders of farm animals and occasionally in pathogenetic syndromes in humans. Both in man and in animals T-2/HT-2 toxins can cause alimentary toxic aleukia. Most controlling government agencies worldwide have regulations or recommendations regarding the amount of T-2/HT-2 allowable in human and animal foodstuffs. Accurate and rapid determination of T-2/HT-2 presence in commodities is of paramount importance.

3. Principle of the Method

The quantitative test is based on the enzyme linked immunosorbent assay principles. The wells of the microtiter strips are coated with T-2 specific antibodies. Toxins are extracted from a ground sample with methanol 70%. Trichothecene standards or samples and T-2-HRP conjugate (detection solution) are added into the coated wells. T-2-HRP conjugate binds to the binding sites of coated antibodies that are not already occupied by T-2 of standards or samples. Any unbound T-2-HRP conjugate of detection solution is removed in a washing step. A chromogen substrate is added to the wells resulting in the progressive development of a blue colored complex with the detection antibody. The color development is then stopped by the addition of acid turning the resultant final product yellow. The measurement is made photometrically at 450 nm and the intensity of the produced colored complex is indirectly proportional to the concentration of T-2/HT-2 present in the samples and standards.

4. Reagents Provided

Bio-Shield T-2/HT-2 ELISA kit contains sufficient reagents and materials for 48/96 measurements (including standard tests).

Reagents (Store at 2-8°C)	Quantity for 48 wells	Quantity for 96 wells	State	Vial cap color
Single-Break Strip Plate	48 wells	96 wells	Ready to use (precoated)	-
Dilution Microwells	48 wells	96 wells	Ready to use (green color)	-
Standards 1-5 (0, 2, 8, 20, and 50ppb of T-2 in organic solution) (correspond to 0, 20, 80, 200 and 500ppb)	5 glass vials (each 1.5ml)	5 glass vials (each 1.5ml)	Ready to use	Black
T-2 Detection Solution	1 plastic vial (12ml)	2 plastic vials(12ml)	Ready to use	Green
Wash Buffer	1 plastic vial (50ml)	1 plastic vial (50ml)	20X Concentrate (dilute in distilled)	White
TMB Substrate	1 plastic vial (6ml)	1 plastic vial (12ml)	Ready to use	Brown
Stop Solution	1 plastic vial (6ml)	1 plastic vial (12ml)	Ready to use	White

5. Materials required but not provided

- A grinder sufficient to render sample to particle size of fine instant coffee
- Balance with 0-50g measuring capability and Graduated cylinder - 100mL
- Methanol (70mL reagent grade per sample) and Distilled or deionized water
- Filter Paper Whatman #1 or equivalent, Filter Funnel and Miscellaneous laboratory plastic or glass tubes 50-125ml
- Vortex mixer and Microtiter plate reader fitted with 450 nm filter
- 100, 200 and 1000µl adjustable single channel micropipettes with disposable tips (a repetitive pipette of 100µl is acceptable for the steps of TMB and Stop Solution)
- 50-300µl multi-channel micropipette with disposable tips and reservoirs

6. Storage Instructions

Store kit reagents between 2 and 8°C (35-46°F). Do not freeze any components provided. Reseal immediately the unused strips of the microtiter plate in the bag together **with the desiccant bag** provided and store at 2-8°C. After use remaining reagents should be returned to cold storage (2-8°C). Expiry of the kit and reagents is stated on the labels respectively and no quality guarantee is accepted after the expiration date. The expiry of the kit components can only be guaranteed if the components are stored properly as well as if the reagent is not contaminated by the first handling, in case of repeated use of one component. Because of the colorless TMB Substrate and standards light sensitivity, avoid the exposure to direct light. Do not interchange individual reagents between kits of different lot numbers.

7. Safety and Precautions for use

- Avoid any skin contact with T-2 standards, Stop Solution (15% H₃PO₄) and TMB (toxic). **Use gloves.** In case of contact, wash thoroughly with water.
- All reagents should be warmed in room temperature before use and covered when not in use. **Use a clean disposable plastic pipette tip for each reagent, in order to avoid cross contamination. When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells.**
- Use a clean plastic container to prepare the wash buffer and all residual washing liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper into the well. Read the absorbance within 60 minutes after completion of the assay.

8. Indication of corruption of kit reagents

- The bluish coloration of the chromogen substrate before the ELISA test.
- A value of less than 0.7 absorbance units (450nm) for the Standard 1 (St1).

9. Sample preparation

9.1 Ground Samples

- Prepare the Extraction Solution (70% Methanol) by adding 30ml of distilled or deionized water to 70ml of methanol (reagent grade) for each sample to be tested.

- The sample must be collected according to established sampling techniques. Grind a representative sample to the particle size of fine instant coffee (50% passes through a 20 mesh screen).
- Weigh out a 20g ground portion of the sample and add 100mL of the Extraction Solvent (70% methanol) and mix in a blender for a minimum of 2 minutes.
- Allow the particulate matter to settle, then filter 5-10ml of the extract through a Whatman #1 filter paper (or equivalent), collect the filtrate and dilute 2 times with deionized water (example: 1ml filtrate + 1ml deionized water). **The ratio of sample to 35% methanol is 1:10 (w/v).**
- Use 100µl of each final diluted filtrate directly in the immunoassay.

NOTE: The extracted sample should have pH value of 6.2-7.5. If the pH is less than 6.2, it should be neutralized using NaOH.

9.2 Milk Samples

- After centrifugation at 3000xg at 4°C for 10 min, remove the upper fat layer, dilute the defatted milk sample 5 times with 35% methanol (**1ml of milk + 4ml of 35% methanol**) and vortex. Use 100µl of each diluted milk sample directly in the immunoassay and divide the final T-2/HT-2 ppm result by 2. The LOD is 2.5ppb and the LOQ is 5ppb.

10. Method Procedure

10.1 Assay Design: Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for standards. Considering that each sample and standard should be tested in duplicate, create a layout. **NOTE:** Do not use more than 48 wells (six strips) in a single experiment.

CAUTION: Use the standards positions in duplicate as the **Example plate** layout below **NECESSARY** and note positions of samples that can be set to all remaining empty wells of layout in duplicate.

	1	2	3	4	5	6	7	8	9	10	11	12
A	St1	St1										
B	St2	St2										
C	St3	St3										
D	St4	St4										
E	St5	St5										
F												
G												
H												

Example plate layout (example for a 5 point standard curve)

10.2 Bring all reagents to room temperature (19-24°C) before use. Remove the **standards** (Standard 1-5) and place two Dilution Microwells (green) in a microwell holder for each Standard and Sample to be tested in duplicate. Place an equal number of Antibody Coated Microtiter Wells in another microwell holder. Immediately reseal the unused strips of the microtiter plate in the bag together with the desiccant bag provided. The samples should be stored in a cool place.

10.3 Add **200µl** of **T-2 Detection Solution** to each Dilution Well.

10.4 Using new pipette tip for each, add 100µl of each Standard (Standard 1 - 5) and prepared sample in duplicate (see Chapter 9) to appropriate Dilution Well containing the Detection Solution. Mix by priming pipetting at least 5 times.

10.5 Using a multichannel pipette, transfer **100µl** of contents from each Dilution Microwell to a corresponding Antibody Coated Microtiter Well. Cover the microwells with the sealing film and incubate at room temperature for **10min**. **NOTE:** During this step, prepare the Wash Buffer 1X working solution (see step 10.6).

10.6 Dilute the 20X solution concentrate 20 fold with distilled water to give a **1X** working solution.

Preparation of Wash Buffer 1X: In case of the occurrence of crystals in the Wash Buffer, the warming by gentle dismantling (using hands) of the crystals is needed. Pour entire content of the solution concentrate (50ml) into a clean 1000ml graduated cylinder, rinse the vial with distilled or deionised water and pour the content again into the cylinder and fill to a final volume of 1000ml with distilled or deionised water. Mix gently to avoid foaming, transferring the final solution from cylinder to a clean bottle and back two times. The clean bottle with **1X Wash Buffer** working solution can be left out of the refrigerator during the method procedure and subsequent be stored 2-8°C for one month.

10.7 Remove the sealing film and wash the plate as follows: Aspirate the liquid from each well into the sink and tap the holder of microwells upside down strongly (four times in a row) on an absorbent paper to insure the complete removal of liquid from the wells. Dispense **300µl** of **Wash Buffer 1X** (see 10.6) into each well with wash bottle or multichannel micropipette using the proper reagent reservoir and shaking the plate manually for a few seconds. Repeat this process for another three times (**total 4 times**). **CAUTION:** It is important to not allow microwells to dry between working steps.

10.8 Aspirate the liquid as described above and add **100µl** per well of **TMB Substrate** (pour 1ml per 8 wells in a reservoir). Cover the microwells with the sealing film, shaking the plate manually for a few seconds and incubate in the dark at room temperature for **5min**.

10.9 Remove the sealing film and add **100µl** per well of the **Stop Solution** to each well (pour 1ml per 8 wells in a reservoir). Mix gently by shaking again the plate manually.

10.10 Measure the absorbance at 450nm. Read the absorbance value of each well (within 60 minutes after the step 10.9) on a spectrophotometer using 450 nm as the primary wavelength and optionally 620nm as the reference wave length (610nm to 650nm is acceptable).

11. Data Analysis

- Automatically

An assigned software, the **Prognosis-Data-Reader**, is available by free download from www.prognosis-biotech.com in order to evaluate the Bio-Shield T-2/HT-2 ELISA kit. The evaluation is carried out by a simple transfer of data values after the measurement.


- Manually

Calculate the average absorbance values for each set of duplicate standards and samples. Ideally duplicates should be within 10% of the mean. Use the following calculation:

$$\frac{\text{Standard or sample absorbance}}{\text{Standard 1 absorbance}} \times 100 = \% \text{ Binding}$$

The standard 1 is equal to 100% and the absorbance values are quoted in percentages. The concentration of Trichothecene (ppb) in each sample is determined by extrapolating OD values against concentrations of Trichothecene in standard solutions using a two phase exponential decay standard curve with logarithmic X axis.

12. Example of Standard Curve (0-500ppb)



13. Immunoassay Specification

13.1 General Specification

- IC50 = 40 - 160 ppb.
- Each standards duplicates mean CV \leq 6%.

13.2 LOD - LOQ - Accuracy

- The LOD of the method is 5ppb.
- The LOQ of the method is 10ppb.
- The recovery of spiked extractions-matrices was 96.4% (CV = 4.1%).
- **Matrices:** **Grains:** Corn, Wheat, Barley, Rice, Oats, Oat Bran, Soy Beans. **Other:** Sesame, Beans, Milk.

13.3 Specificity

The cross-reaction of the anti-T-2 antibody with HT-2, T-2 Triol and T-2 Tetraol is 80, 3.7 and <0.1% respectively.

14. Performance Evaluation

14.1 Reference Materials

Reference material	Lot number	Certified value ($\mu\text{g}/\text{kg}$)	Range for $ z \leq 2$ ($\mu\text{g}/\text{kg}$)	Result ($\mu\text{g}/\text{kg}$)
BIPEA Corn Mycotoxins 0031 November 2016	3-1167	82	41	76.9

14.2 Proficiency Tests

Test	Assigned Value ($\mu\text{g}/\text{kg}$)	Result ($\mu\text{g}/\text{kg}$)	Z-score
BIPEA 31b 155 Mycotoxins - Ground Rye Sample 01 - 5131 June 2017	163	181	0.42
BIPEA 31c 9 Mycotoxins - Baby Food - 09 - 3931 - June 2017	37	45	0.73

15. Method Summary

Total procedure time (after samples and reagents preparation): 15min.

Mix 200µl of the Detection Solution with 100µl of the samples and standards in the Dilution Microwells



Transfer 100µl from each well of the Dilution Microwells into the Antibody Coated Microwells



Incubate 10min at room temperature



Wash four times



Add 100µl of TMB Substrate



Let the color develop for 5min in the dark at room temperature



Add 100µl Stop Solution



Read Absorbance at 450nm within 60min

All immune assays supplied by ProGnosis Biotech Ltd, are warranted to meet or exceed our published specification when used under normal conditions in your laboratory. If the product fails during the stated period, a replacement product will be issued.

ProGnosis Biotech makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. There is no warranty of merchantability of this product, or of the fitness of the product for any purpose. ProGnosis Biotech shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from the use of this product. This method is considered to be a screening method, before a legal action, samples detected as positives must be confirmed with a confirmation method. This product is meant to be used only For Research or Manufacturing use and by qualified technicians.



FOLLOW @PROGNOSIS_



JOIN US ON
FACEBOOK



PROGNOSIS

BIOTECH

ProGnosis Biotech Ltd

R&D, In vitro diagnostics, Biotechnology Marketplace

Iroon Polytechniou 71

Larissa, Greece, 41222

tel: 2410 623922 / 2310 952738

fax: 700 700 6262

WEB: WWW.PROGNOSIS-BIOTECH.COM

E-MAIL: INFO@PROGNOSIS-BIOTECH.COM